



Insert the ground cord's banana plug into the ground jack located "J" at the bottom of the tester. Connect ground cord's ring terminal to equipment ground. This connection will remove any static charge from the user before the test.

NOTE:

Failure to correctly ground the Combo Tester X3 may result in damage not covered under warranty!



Vor Inbetriebnahme des Gerätes muss zwingend die Erdungsleitung an die Erdungsbuchse „J“ angeschlossen werden.

Achtung:

Ansonsten besteht die Gefahr eines Defektes, der nicht in den Garantieleistungen enthalten ist!